

Laboratory and Machinery Equipment

No	Tender No	Specification(DMR)	Specification(Co.,Ltd)	Unit	Co.,Ltd	Remark
1	E-1	Vitek Densicheck (Biomerieux) (Automated McFarland measurement for Vitek) List of Components,(1) Vitek Densicheck Display Base, (Product code 422220), (2) Densicheck Display POD (Product code 423000), (3)Densicheck MC FARLAND Reference Kit (Productcode 422219) Voltage-100-240VAC, Frequency - 50-60 Hz, Current-1A, Power- 30 Watts	Vitek Densicheck (Biomerieux) (Automated McFarland measurement for Vitek) List of Components , (1) Vitek Densicheck Display Base(Product code 422220) (2) Densicheck Display POD(Product code 423000), (3) Densicheck MC FARLAND Reference Kit (Product code 422219) Voltage - 100-240VAC, Frequency - 50-60 Hz, curent-1A, Power-30 WattsCountry of Origin; Biomerieux/USA	1 Set	Okkar Thiri	
2	E-2	Electrocardiograph,Specification: ECG 3 Channel Machine.Main Unit with 5" TFT color LCD screen, for Simultaneous 12 Lead ECG Acquisition.HR detection alarm and pacemaker detection supported. Battery Operation Time: ≥ 180 minutes Resting ECG recording format: 1, 3, Average + Rhythm	Electrocardiograph Three-channel DigitalElectrocardiograph Country of Origin; iE 300 Biocare/China	1 No	Okkar Thiri	

Non-Control Items

No	Tender No	Specification(DMR)	Specification(Co.,Ltd)	Unit	Co.,Ltd	Remark
1	NC-1	Vitek 2 GN cards, (Product code 21341), (Biomerieux), (20 cards/pkt)	Vitek 2 GN Cards, (Product Code 21341), (Biomerieux) (20 cards/pkt),Biomerieux USA	5 boxes	Okkar Thiri	
2	NC-2	Vitek 2 GN AST cards N 405, (Product code 423864), (Biomerieux), (20 cards/pkt)	Vitek 2 GN AST cards N 405, (Product code 423864), (Biomerieux), (20 cards/pkt)	5 boxes	Okkar Thiri	
3	NC-3	Vitek 2 GP cards , (Product code 21342), (Biomerieux), (20 cards/pkt)	Vitek 2 GP Cards(Product Code 21342), Biomerieux) (20 cards/pkt) Biomerieux USA	5 boxes	Okkar Thiri	
4	NC-4	Vitek 2 GP 67 AST cards (Product code 22226), (Biomerieux) (20 cards/pkt)	Vitek 2 GP 67 AST Cards(Product Code 22226), (Biomerieux) (20 cards/pkt) Biomerieux USA	5 boxes	Okkar Thiri	
5	NC-5	Unsensitized tubes, (Product code 69285), (Biomerieux), (2000 tubes/pkt)	Unsensitized tubes(Product code 69285), (Biomerieux) (2000 tubes/pkt) Biomerieux USA	1 box	Okkar Thiri	
6	NC-6	3500XL Genetic Analyzer Capillary Array, 24-Cap, 50cm, ABI/ Thermo Scientific, Cat number: 4404689	3500XL Genetic Analyzer Capillary Array, 24-Cap, 50cm Model:4404689, Brand:ABI/ ThermoFisher Scientific, COO:USA	1 No	Sciencewerke (Myanmar)	
7	NC-8	Rotavirus Rapid Test Kit (Rotavirus antigen detection immunochromato- graphic test, Cassette Type)	Rotavirus Rapid Test kit (Rotavirus antigen detection immunichromatographic test, Cassette Type) LumiQuick Diagnostic, Inc., USA	500 ea	Focus Health	
8	NC-10	Online UPS, Specification, Capacity : 40kVA,	“ABLEREX” KRONOS Series 40kVA Online UPS	1 No	Maha Power	

		<p>Input Voltage : 380/400/415V, 3 Phase, Input frequency : 40~70Hz, Output Voltage : 380/400/415V, 3 Phase, Output frequency : 50/60Hz, Efficiency : up to 95%, Back up time : 10 min for 70% , nominal load, Operating temp : 0-40°C</p>	<p>System, Rated Input Voltage : 380/400/415, 3 Phase +N, Voltage tolerance:± 20%@100% load, Input frequency : 40~70Hz, Input power Factory/ THDI: 0.99/< 3% @ 100% Load, Output Voltage : 380/400/415V, 3 Phase +N, Output frequency : 50/60Hz, Output wave form: Pure sine wave, Output Overload: 125% for 10 mins,150% for 1 min, Operating temp : 0°C ~ 40°C, Efficiency : up to 95%, Battery System- External complete with open Rack , Battery Brand: Ener Rocket (Korea), Battery Capacity :12V,42Ah, Battery Quantity:36 Blocks, Backup time : 10 min @ 70% , nominal load, UPS Dimension (W x D x H): 260 x 890 x 850 (mm),UPS System, COO: ABLEREX(Italy), Manufacturer: Manufactured in China, Model of UPS: KRONOS 40S- FOS. Battery System, Type of DC Cable: H07RNF Flexible DC Cable(Made in Italy/Spain),Type of Battery Breaker: ABB Brand DC Breaker (Made in Italy), COO of Battery: Ener Rocket (Korea). Manufacturer of Battery: Vietnam</p>		<p>Myanmar</p>	
--	--	---	---	--	----------------	--

Office Equipment

No	Tender No	Specification(DMR)	Specification(Co.,Ltd)	Unit	Co.,Ltd	Remark
1	OE-1	Projector, LCD Projector WXGA (1280 x 800) 5500 ANSI Lumen 5000 Hour Normal Lamp Life 16000:1 Contrast 1.6xZoom 30 to 300 inch Screen Size 10W Speaker Direct Power Off Easy Remote Operation Daylight View HDMIx2 USB Memory Viewer RJ45 Vertical Lens Shift Horizontal Lens Shift	Projector, LCD Projector WXGA (1280 x 800) 5500 ANSI Lumen 5000 Hour Normal Lamp Life 16000:1 Contrast 1.6xZoom 30 to 300 inch Screen Size 10W Speaker Direct Power Off Easy Remote Operation Daylight View HDMIx2 USB Memory Viewer RJ45 Vertical Lens Shift Horizontal Lens Shift, Model: EH 500, Brand: Optoma, Country Of Origin: China	1 No	Concordia International	
2	OE-2	Paper Cutter, 1710 W Trimmer, Sheet Capacity:10 sheet Cutting Length: A3-432 mm (17") Edge Glow LED illumination Durable Solid Wood Surface A Safety Latch AAA Battery	Paper Cutter, W Trimmer, Sheet Capacity : 10 Sheet Cutting, Length : A3--432 mm (17")EDGE Glow LED, illumination Durable Solid Wood Surface A Safety, Latch AAA Battery, Warranty : 1 Year Parts and Lifetime in service warranty. Brand : Rexel, Model : 1710W, Country of Origin : China (Made in China)	3 No	Gandamar Business Solution	
3	OE-3	Binding Machine, Plastic Comb Binding Machine, Binding Style: Plastic Comb Bind Document Size : A4 Punch Capacity:20 Sheet, No. of Punch Holes:21 Bind Capacity: 330 Sheets, Dimension LWH: 395x340x130 mm, Weight: 5.3 kg	Binding Machine, Plastic Comb Binding, Machine, Binding Style : Plastic , Comb Bind Document Size : A4 Punch Capacity : 20, Sheet, No. Of Punch Holes:21 Bind Capacity : 330, Sheets. Dimension LWH 395 x 340 x 130mm Weight: 5.3 Kg, Warranty : 1 Year Parts and Lifetime in service warranty. Brand : GBC, Model : CombBind200, Country of Origin : Portugal (Made in China)	2 No	Gandamar Business Solution	

--	--	--	--	--	--	--

**သုတေသနဓါတ်ခွဲခန်းသုံးစက်နှင့်စက်ပစ္စည်းများ၊ ဓါတ်ခွဲခန်းသုံးပစ္စည်းများနှင့် ရုံးသုံးစက်ပစ္စည်းများ
ဝယ်ယူရန်အတွက် တင်ဒါအောင်စာရင်း ထုတ်ပြန် ကြေညာခြင်း**

၂၀၂၄-၂၀၂၅ခု ဘဏ္ဍာရေးနှစ်တွင် ဆေးသုတေသနဦးစီးဌာန(ရုံးချုပ်) နှင့် (ပြင်ဦးလွင်ရုံးခွဲ)၌ အသုံးပြုရန်ဝယ်ယူမည့် သုတေသနဓါတ်ခွဲခန်းသုံးစက်နှင့်စက်ပစ္စည်းများ၊ ဓါတ်ခွဲခန်းသုံးပစ္စည်းများနှင့် ရုံးသုံးစက်ပစ္စည်းများ၏ တင်ဒါအောင်စာရင်းကို ၃၁ - ၅ - ၂၀၂၄ ရက်နေ့တွင် ဆေးသုတေသနဦးစီးဌာန (ရုံးချုပ်)၌ ထုတ်ပြန်ကြေညာလိုက်သည်။

တင်ဒါအောင်မြင်သောကုမ္ပဏီများသည် ဆေးသုတေသနဦးစီးဌာန(ရုံးချုပ်)သို့ အမြန်ဆက်သွယ် ဆောင်ရွက်ရန်နှင့် တင်ဒါမအောင်မြင်သောကုမ္ပဏီများသည် တင်ဒါစပေါ်ငွေ(၁%)အား ပြန်လည် ထုတ်ယူနိုင်ရေးအတွက် ဆေးသုတေသနဦးစီးဌာန(ရုံးချုပ်)သို့ ပြန်လည်တင်ပြလျှောက်ထားရန် ဖြစ်ပါသည်။

တင်ဒါလက်ခံရေးနှင့်စိစစ်ရေးကော်မတီ
ဆေးသုတေသနဦးစီးဌာန